

PAPER • OPEN ACCESS

Background Radiation Dose in Selected X-Ray Facilities in Southwest Nigeria

To cite this article: J. A. Achuka *et al* 2019 *J. Phys.: Conf. Ser.* **1299** 012103

View the [article online](#) for updates and enhancements.



IOP | ebooks™

Bringing together innovative digital publishing with leading authors from the global scientific community.

Start exploring the collection—download the first chapter of every title for free.

Background Radiation Dose in Selected X-Ray Facilities in Southwest Nigeria

^{a*}J. A. Achuka, ^aM. R. Usikalu, ^bM. A. Aweda, ^aC. A. Onumejor and ^aI. O. Babarimisa

^aDepartment of Physics, Covenant University Ota, Ogun State, Nigeria

^bDepartment of Radiation Biology, Radiotherapy and Radiodiagnosis, College of Medicine, Lagos University Teaching Hospital, Idi-Araba, Lagos, Nigeria

justina.achuka@covenantuniversity.edu.ng

Abstract. Radiation exposure monitoring is essential to ensure that dose limits are not exceeded. The goal of this study is to assess the level of radiation exposure from radiography facilities in the study environment in order to promote radiation safety. Digilert 200 was used to determine the level of exposure in and around radiography facilities in five X-ray diagnostic centres in southwest Nigeria. The mean background exposure in centres A, B, C, D and E were $0.137 \mu\text{Sv/h}$, $0.170 \mu\text{Sv/h}$, $0.164 \mu\text{Sv/h}$, $0.183 \mu\text{Sv/h}$ and $0.148 \mu\text{Sv/h}$ respectively. The cumulative mean exposure in a year for centres A, B, C, D and E were 0.961 mSv/y , 1.193 mSv/y , 1.146 mSv/y , 1.281 mSv/y and 1.034 mSv/y respectively. The background radiation dose from the exposure level in all the centres exceeded the recommended limit but for centre A. High quality standard lead shielding and periodic radiation protection monitoring should be employed in centres with high radiation exposure.

Keywords: Background radiation, radiography facilities, exposure level, radiation safety

1. Introduction

Radiation exposure from diagnostic radiology is fast increasing and it has become a growing concern [1-3]. Dramatic explosion in imaging technology and the extensive use of these modalities are responsible for the increased in background radiation exposure. All X-ray tubes are known to have some radiation leakage [4]. This warrants the measurement of the background radiation to ensure the exposure is within the acceptable limits. Based on this finding, it is recommended that X-ray diagnostic facilities should be subjected to standard safety operations considering the risks of ionizing radiation [5]. Contrarily, several X-ray diagnostic facilities are not radiologically safe. Poor radiation protection measures in diagnostic radiology have been reported in literature [4, 6-8]. Report of [4] showed that the fluoroscopy and CT rooms were not adequately lead lined and the radiation exposure exceeded the reference limit in Gaza. However, the radiography and mammography rooms exposure were within the permissible limit. Lack of radiological protection act and monitoring has been reported in Nepal [7]. Another study reported the absence of radiation protection tools and equipment in Saudi Arabia [8]. This further confirms the poor radiation protection measures in some X-ray unit.

Association between radiation exposure and diagnostic X-ray imaging has been reported [8]. The biologic effect of diagnostic X-ray exposure is largely stochastic effects [9, 10]. The major stochastic



effects of concern for diagnostic X-rays are cancer induction and genetic effects [9] Though there have been several arguments on stochastic effect at low dose of exposure yet, recent studies has further confirmed that low radiation exposure can induce DNA damage regardless of the dose [11-13]. Literature has also reported induce DNA damage, chromosomal aberration, genotoxic effects and other radiation hazards in tissue for non-ionizing radiation [14-16]. Hence, there is need for safety. Therefore, this study goal is to assess the background radiation exposure in and around selected radiography facilities in southwest Nigeria in order promote safety and minimize health hazards associated with radiation exposures.

2. Materials and Methods

The study was conducted in five (5) secondary/tertiary healthcare institutions in southwest Nigeria with an average of 400 radiography examinations per day. The scatter radiation was determined using the pre-calibrated Digilert 200 survey meter. The survey meter was calibrated to standard at the Secondary Standard Dosimetry Laboratories (SSDL) of National Institution for Radiation Protection and Research, University of Ibadan. Determination was made at distance of 1 m from the back, front, left and right of the X-ray system. Also, measurements were made at the console, patient waiting rooms, reception, dark room and corridors. At each location measurements were made at a distance of 1 m above the ground with an average of 5 radiation counts recorded. The mean values were obtained in mR/hr and later converted to $\mu\text{Sv/h}$.

3. Results and Discussion

The results of the mean background radiation in and around each radiography unit in the study centres are as presented in Table 1. The estimated cumulative mean exposure in centres A, B, C, D and E were $0.137 \mu\text{Sv/h}$, $0.170 \mu\text{Sv/h}$, $0.164 \mu\text{Sv/h}$, $0.183 \mu\text{Sv/h}$ and $0.148 \mu\text{Sv/h}$ for centres respectively (Figure 1). The cumulative mean exposure in a year is as shown in Figure 2. The estimated background radiation exposure was above the recommended limit of 1 mSv/y in all the centres but for centre A.

Radiation monitoring is essential to ensure that dose limit is not exceeded [17-18]. The results from this study showed that the cumulative background radiation exceeded the recommended limit of 1 mSv/y in all the centres but for centre A. The report in this study is in consonance with study conducted in Gaza and Saudi Arabia [4, 8]. The high exposure level recorded in the study centres can be attributed to poor lead shielding, structural fault and lack of radiation monitoring. Research has shown that any defects in the structural design of the imaging room can lead to higher exposure levels [7, 19].

Centre A has low radiation exposure compared to other centres because of high quality of lead shielding although it has structural defect. The exposure level in the console has the least value as compared to others locations. The value of $0.090 \mu\text{Sv/h}$ and $0.110 \mu\text{Sv/h}$ was recorded against the corridor and reception respectively in centre E. This might be due to the fact that the corridor and reception are farther away from the scanning room. This is envisaged as X-rays obeys the inverse square law. Centre D has the highest value ($0.183 \mu\text{Sv/h}$). This might be due to the fact that the radiography facility is old and lack adequate maintenance. The lead shielding is very poor and might have expired. The result from this study further confirms the literature report of poor radiological protection in some medical X-ray diagnostic centres.

The challenge of radiation protection in diagnostic radiology can be attributed to unavailability of devices for the assessment of radiation exposure, ignorance of health implications of low dose exposure, poor enforcement of radiation protection policy, insufficient skilled radiological personnel, and engagement of unskilled personnel in radiological department among others. Improvement of radiation protection measures in the study centres is therefore important to enhance safety.

Table 1: Background radiation ($\mu\text{Sv}/\text{h}$) around radiography equipment in the study centres

Centre	X-Ray room	Console	Changing room 1	Changing room 2	Processor room	Reception	Corridor	Mean
A	0.157	0.110	0.157	0.157	0.150	0.115	0.115	0.137
B	0.163	0.158	0.130	0.120	0.140	0.240	0.240	0.170
C	0.217	0.177	0.158	0.153	0.160	0.140	0.140	0.164
D	0.210	0.197	0.210	0.210	0.180	0.130	0.143	0.183
E	0.190	0.125	0.190	0.000	0.180	0.110	0.090	0.148




Figure 1: Mean background radiation in the study centres




Figure 2: Cumulative background radiation per year in the study centres

4. Conclusion

Increased knowledge and training on radiological protection enhance safety and minimize health hazards associated with radiation exposures. It is evident that the study centres lack adequate radiation protection measures. As the radiation exposure level in most centres were above the recommended limit. It is necessary that centres with high exposure should seek professional advice to change their lead shielding to high quality standard. Also, periodic radiation protection monitoring should be constituted.

Acknowledgements

Special thanks to Covenant University Centre for Research, Innovation and Discovery for the financial support and all the healthcare institutions used.

References

- [1] Boone, J. M., Hendee, W. R. McNitty-Gray, M. F. and Seltzer, S. E. (2012). Radiation exposure from CT scans: How to close our knowledge gaps, monitor and safeguard exposure. Proceedings and Recommendations of the Radiation Dose Summit, Sponsored by NIBIB, February 24-25, 2011. *Radiology*, 265 (2): 544-554.
- [2] Kathmann, W. (2017). Optimization of radiation dose performance. *Diagnostic Imaging*, 33 (2): 48-50.
- [3] Scharf, M., Brendel, S., Melzer, K., Hentschke, C., May, M., Uder, M. and Lell, M.M. (2017). Image quality, diagnostic accuracy, and potential for radiation dose reduction in thoraco-abdominal CT using sonogram affirmed iterative reconstruction algorithm (SAFIRE) technique in a longitudinal study. *PLOS ONE*, 12 (7): e0180302.
- [4] Abu-Zer, S. S., Khadoura, K. J., Yassin, S. S., Agha, M. R. (2016). Ionizing radiation leakage in radiodiagnostic centers at Gaza Strip hospitals, Palestine. *Asian Review of Environmental and Earth Sciences*, 3 (1): 18-26.
- [5] Panicker, T.M. R., Tina-Angelina, J.T., Korath, M.K., Mohandas, K. and Jagadeesan, K. (2013). Entrance skin dose estimation in X-ray lumbar spine lateral procedure: Conventional vs digital X-ray units: A pilot study. *JIMSA*, 26 (4): 219-220.

- [6] Bhatt, C. R., Widmark, A., Shrestha, S. L., Khanal, T., Ween, B. (2012). Occupational radiation exposure in health care facilities. *Kathmandu University Medical Journal*, 39 (3): 48-51.
- [7] Adhikari, K. P., Jha, L. N., Galan, M. P. (2012). Status of radiation protection at different hospitals in Nepal. *Journal of Medical Physics*, 37 (4): 240-244.
- [8] Salama, K. F., Al-Obireed, A., Al-Bagawi, M., Al-Sufayan, Y., Al-Serheed, M. (2016). Assessment of occupational radiation exposure among medical staff in healthcare facilities in the eastern province, Kingdom of Saudi Arabia. *Indian Journal of Occupational and Environmental Medicine*, 20 (1): 21-25.
- [9] International Atomic Energy Agency (IAEA, 2014). *Diagnostic radiology Physics: A handbook for teachers and students*. IAEA, Vienna. ISBN 978-92-131010-1, 1-710.
- [10] Nuclear Regulatory Commission (NRC, 2014). *Fact sheet on biological effect of radiation*. U.S NRC, accessed: 14/03/2018.
- [11] Kawamura, K., Qi, F. and Kobayashi, J. (2018). Potential relationship between the biological effects of low dose irradiation and mitochondrial ROS production, *Journal of Radiation Research*, 59(2): 91-97.
- [12] Nagashima, H., Shiraiishi, K., Ohkawa, S., Sakamoto, Y., Komatsu, K., Matsuura, S., Tachibana, A. and Tauchi, H. (2018). Induction of somatic mutations by low dose X-rays: The challenge in recognizing radiation induced events. *Journal of Radiation Research*, 59 (2): 11-17.
- [13] Aweda M. A., Usikalu M. R., Ding N, Wan J. H, Zhu J (2010). Genetoxic effects of 2.45 GHz microwave exposure on different cells of Sprague Dawley rats *International Journal Genetics and Molecular Biology* 2(9): 189-197.
- [14] Usikalu M.R., Aweda M.A., Alimba C.G., Achuka J.A. (2016). Chromosomal aberration after exposure to 2.45 Ghz microwave radiation. *Research Journal of Applied Sciences*, 11(5): 232-234.
- [15] Usikalu M.R., Obembe O.O., Akinyemi M.L., Zhu J. (2013). Short duration exposure to 2.45 Ghz microwave radiation induces DNA damage in Sprague Dawley rat's reproduction systems, *African Journal of Biotechnology*, 12 (2): 115-122.
- [16] Achuka, J.A., Aweda, M.A., Usikalu, M.R. (2018). Cancer risks from head radiography procedures. *Earth and Environment Science*, 173:012038. doi:10.1088/1755-1315/173/1/012038.
- [17] International Commission on Radiological Protection (ICRP, 2007). *Radiological protection in medicine* Publication 105. *Annals of the ICRP*.
- [18] United States Environmental Protection Agency (USEPA, 2014). *Radiation protection guidance for diagnostic and interventional X-ray procedures*. Federal Guidance Report No 14, EPA-402-R-10003, Washington D.C.
- [19] Dehaghi, B. F., Ghavamabadiz, L. I., Bozar M., Mohamadi, A., Angali, K. A. (2017). Evaluation of X-ray radiation levels in radiology departments of two educational hospitals in Ahvaz, Iran. *Iran Journal of Medical Physics*, 14 (2): 87-91.